

u^{*b*}

b

**UNIVERSITY
OF BERN**

u^b X-ray microtomography

485018-HS2024-0: Advanced Course II Ultraprecision Engineering

David Haberthür

Institute of Anatomy, September 25, 2024

u^b Grüessech mitenang!

- David Haberthür
 - Physicist by trade
 - PhD in high resolution imaging of the lung, Institute of Anatomy, University of Bern, Switzerland
 - Post-Doc I: TOMCAT, Swiss Light Source, Paul Scherrer Institute, Switzerland
 - Post-Doc II: μ CT group, Institute of Anatomy, University of Bern, Switzerland

u^b


Grüessech from the μ CT group



David.Haberthuer@unibe.ch Ruslan.Hlushchuk@unibe.ch Oleksiy.Khoma@unibe.ch

μ CT-group

- microangioCT [1]
 - Angiogenesis: heart, musculature [2] and bones
 - Vasculature: (mouse) brain [3], (human) nerve scaffolds [4], (human) skin flaps [5] and tumors
- Zebrafish musculature and gills [6]
- (Lung) tumor detection and metastasis classification [7]
- Collaborations with museums [8] and scientist at UniBe [9] to scan a wide range of specimens
- Automate *all* the things! [10, 11]



bruker.com/skyscan1272

Contents

Overview & Imaging methods

Tomography

History

Interaction of x-rays with matter

Tomography today

A scan, from *getting started* to *nice image*

Examples

A study about teeth

Overview

Materials & Methods

Results

Metal foam analysis

A study on fish

- Dense and/or non-transparent samples
- Calibrated & isotropic 3D images at micron resolutions
- Covers a very large range of sample sizes
- Gives information at different length scales
- Nondestructive imaging, thus compatible with routine sample preparation.
Enables correlative imaging pipelines, scanning of museum & collection material

u^b

Biomedical imaging

- Medical research
- Non-destructive insights into the samples
- (Small) Biological samples




w.wiki/7g4 © ⓘ


u^b

Biomedical imaging

- Medical research
- Non-destructive
- (Small) Bio




u^b Wavelengths & Scales



w.wiki/7fz © ⓘ ⓘ

u^b Wavelengths & Scales




Stefan Tschanz, with permission

Imaging methods

- Light (sheet) microscopy: see lecture of Nadia Mercader Huber
- X-ray imaging
- Electron microscopy
 - *Analytical electron microscopy* by Dimitri
 - *SEM Grundlagen* by Sabine Kässmeyer and Ivana Jaric
 - *Cryoelectron Microscopy & Serial Block Face SEM* by Ioan

u^b

CT-Scanner



youtu.be/2CWpZKuy-NE

CT History


- 1895: Wilhelm Conrad Röntgen discovers X-rays



w.wiki/BHAN ©

CT History

- 1895: Wilhelm Conrad Röntgen discovers X-rays
- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [12]
- 1976: Hounsfield worked on first clinical scanner [13]





From [14], Figure 5

CT History

- 1895: Wilhelm Conrad Röntgen discovers X-rays
- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [12]
- 1976: Hounsfield worked on first clinical scanner [13]
- Nobel Prize in 1979, jointly for Allan Cormack and Godfrey Hounsfield


U.S. Patent Feb. 24, 1976 Sheet 1 of 2 3,940,625



From [US3940625A], p. 2

CT History


- 1895: Wilhelm Conrad Röntgen discovers X-rays
- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [12]
- 1976: Hounsfield worked on first clinical scanner [13]
- Nobel Prize in 1979, jointly for Allan Cormack and Godfrey Hounsfield
- CT scanner generations
 - First generation



From [15], Figure 1.12

CT History


- 1895: Wilhelm Conrad Röntgen discovers X-rays
- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [12]
- 1976: Hounsfield worked on first clinical scanner [13]
- Nobel Prize in 1979, jointly for Allan Cormack and Godfrey Hounsfield
- CT scanner generations
 - First generation
 - Second generation



From [15], Figure 1.13

CT History

- 1895: Wilhelm Conrad Röntgen discovers X-rays
- 1963: Cormack used a collimated ^{60}Co source and a Geiger counter as a detector [12]
- 1976: Hounsfield worked on first clinical scanner [13]
- Nobel Prize in 1979, jointly for Allan Cormack and Godfrey Hounsfield
- CT scanner generations
 - First generation
 - Second generation
 - Third generation



From [15], Figure 1.14

μ CT History I

- X-ray computed tomography began to replace analog focal plane tomography in the early 1970s [**Lin2019**]
- Non-medical use in the late 1970s, for detection of internal defects in fabricated parts and equipment
- Lee Feldkamp [16] developed an early laboratory microCT system by assembling a micro-focus cone beam x-ray source, specimen holder and staggers, and an image intensifier at Ford Motor Company's Scientific Research Laboratory to nondestructively detect damage in ceramic manufactured automobile parts
- Feldkamp met with scientists at Henry Ford Hospital and University of Michigan interested in understanding the relationship between the microstructure and biomechanical function of trabecular bone to study osteoporotic fractures [**Feldkamp1983**]

μ CT History II

- μ CT was first reported in the 1980s, for scanning gemstones
- Early 1990s: Manufacturers like SkyScan and Scanco Medical made μ CT systems commercially available
- Today: Nondestructive imaging for quantifying the (micro)structure of (organic) materials
 - Mineralized bone tissue and the relationships between the mechanical behavior of bone to its structural and compositional properties
 - Teeth and their internal details
 - Tissues, small animals, and medical devices like stents and implants
 - Soft tissues and vasculature using radio-opaque contrast agents
 - Characterization of anatomical details in high resolution
- ≈ 2500 μ CT systems are in use worldwide with over 1000 publications annually

X-ray interaction

- “X-rays interact with tissue in 2 main ways: photoelectric effect and Compton scatter. To a first approximation, the photoelectric effect contributes to contrast while the Compton effect contributes to noise. Both contribute to dose.” ([19])
 - Photoelectric absorption (τ) is strongly dependent on the atomic number Z of the absorbing material: $\tau \propto \frac{Z^4}{E^{3.5}}$
 - Compton scattering is one of the principle forms of photon interaction and is directly proportional to the (electron & physical) density of the material. It does *not* depend on the atomic number: $\lambda' - \lambda = \frac{h}{m_e c} (1 - \cos \theta)$
- Lowering x-ray energy increases contrast
- X-ray penetration decreases exponentially with sample thickness [20, i. e. Beer-Lamberts law]: $I(t) = I_0 e^{-\alpha z}$


Composition of biological tissues

Tissue: content by mass percentage


Element Atomic number	H 1	C 6	N 7	O 8	Na 11	P 15	S 16	Cl 17	K 19	Ca 20
Fat	11.4	59.8	0.7	27.8	0.1		0.1	0.1		
Water	11.2			88.8						
Blood	10.2	11	3.3	74.5	0.1	0.1	0.2	0.3	0.2	
Liver	10.2	13.9	3	71.6	0.3	0.2	0.3	0.2	0.3	
Brain	10.7	14.5	2.2	71.2	0.2	0.4	0.2	0.3	0.3	
Bone	3.4	15.5	4.2	43.5	0.1	10.3	0.3			22.5

u^b

Why μ CT?




From [21], Subject C3L-02465



u^b

Why μ CT?




From [21], Subject C3L-02465




u^b

Why μ CT?




From [21], Subject C3L-02465




u^b

Why μ CT?



From [21], Subject C3L-02465




u^b

Why μ CT?



From [21], Subject C3L-02465






Maximum intensity projection

Projections

 u^b UNIVERSITÄT
DUISBURG
ESSEN

Reducing the dimensions of a dataset. For example projecting a volume (3D) to a surface by taking the maximum value across planes for each pixel.



u^b

Machinery

- Hospital CT
 - Voxel size around 0.5 mm
- Lab/Desktop CT
 - Voxel size around 7 μm (*in vivo*)
 - Voxel size around 0.5 μm (*ex vivo*)
- Synchrotron CT
 - Voxel size down to 160 nm



flic.kr/p/D4rbom © ⓘ ⓘ ⓘ

u^b

Machinery


- Hospital CT
 - Voxel size around 0.5 mm
- Lab/Desktop CT
 - Voxel size around 7 μm (*in vivo*)
 - Voxel size around 0.5 μm (*ex vivo*)
- Synchrotron CT
 - Voxel size down to 160 nm



flic.kr/p/fpTrGu © ⓘ ⓘ

Machinery

- Hospital CT
 - Voxel size around 0.5 mm
- Lab/Desktop CT
 - Voxel size around 7 μm (*in vivo*)
 - Voxel size around 0.5 μm (*ex vivo*)
- Synchrotron CT
 - Voxel size down to 160 nm



bruker.com/skyscan1272

Machinery

- Hospital CT
 - Voxel size around 0.5 mm
- Lab/Desktop CT
 - Voxel size around $7\text{ }\mu\text{m}$ (*in vivo*)
 - Voxel size around $0.5\text{ }\mu\text{m}$ (*ex vivo*)
- Synchrotron CT
 - Voxel size down to 160 nm




flic.kr/p/7Xhk2Y © ⓘ

What is happening?

No matter what kind of machine, the basic principle is always

- an x-ray source
- a sample
- a detector




w.wiki/7g3 © ⓘ ⓘ

u^b

Machinery


u^b

Examples




u^b

Examples



u^b


Examples



From [8], *Diancta phoenix*

u^b

Examples



u^b

Examples




u^b

Preparation

- Study design
- Sample preparation


Projections

X-ray generation and contrast



Contrast is given by
absorption of intensity I


Note that contrast is negative
X-ray shadowgraphy
is a bright field technique



Contrast, Magnification and Resolution—Laws of Physics for Microscopists (1, 2022) by Martin Frenz, Slide 21


u^b

Projection acquisition



u^b


Projections




Projections

- A (micro-focus) x-ray source illuminates the object
- The x-rays penetrate the sample and are attenuated
- A scintillator converts the x-rays to visible light
- A (planar) x-ray detector collects (magnified) projection images.
- The projections are recorded on disk

Reconstructions




Reconstructions



u^b

Reconstructions



Reconstructions

- Based on hundreds of angular views acquired while the object rotates, a computer synthesizes a stack of virtual cross section slices through the object.
- Radon Transformation
- Filtered back projection
- Cone beam reconstruction [16]
- Corrections (beam hardening, etc.)
- Writing to stack

u^b

Visualization



Visualization

- Based the on reconstructions, a computer synthesizes a three-dimensional view of the scanned sample

What to use?

- ImageJ/Fiji [22]
- Also see *Fundamentals of Digital Image Processing* by Guillaume Witz
- Reproducible research
 -  in Jupyter [23]
 - **git**
 - Script all your things!
 - Data repositories; i. e. sharing is caring!

u^b

Quantitative data

- Pretty images are nice to have, but science is quantitative data
- Segmentation
- Characterization


u^b

Internal morphology of human teeth

Collaboration with zmk bern – Zahnmedizinische Kliniken

– Num'

– S



– Analytic

u^b


How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!



How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!



bruker.com/skyscan1272

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!

```
Scanner=SkyScan1272
Instrument S/N=15G09089-B
Software Version=1.1.19
Filename Prefix=Tooth045~00
Number Of Files= 482
Number Of Rows= 1092
Number Of Columns= 1632
Source Voltage (kV)= 80
Source Current (uA)= 125
Image Pixel Size (um)=9.999986
Exposure (ms)=950
Rotation Step (deg)=0.400
Frame Averaging=ON (3)
Filter=Al 1mm
Study Date and Time=02 Jul 2020
08h:23m:34s
Scan duration=0h:39m:51s
```

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!


Sample changer on the SkyScan 1272

In total:

- 13 days of *continuous* μ CT scanning
- 819 GB of raw data
230 648 TIFF projections
- 326 GB data as input for analysis
282 062 PNG reconstructions

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!




From [25], Fig. 2

u^b

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!



From [27], Fig. 1

u^b

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!



gph.is/2nqkple

How?

- 104 extracted human permanent mandibular canines
- μ CT imaging
- Root canal configuration, according to Briseño-Marroquín et al. [25]
- *Reproducible* analysis [26], e. g. you can click a button to double-check or recalculate the results yourself!


The screenshot shows a GitHub repository interface. At the top, there are tabs for 'master', '1 branch', and '1 tag'. Below these are buttons for 'Go to file', 'Add file', and 'Code'. The repository name is 'habib' and the file being viewed is 'Update actions file'. The commit history shows 'ds59125' from '20 days ago' with '87 commits'. A table lists the repository files:

File	Description	Last Commit
.github/workflows	Update actions file	20 days ago
.gitignore	Only 'mode' changes	2 months ago
DownloadFromOSF.ipynb	Clean run of download script	22 days ago
README.md	Typo in Binder badge & link to full repo on Binder	22 days ago
Tooth.Border.jpg	Only 'mode' changes	2 months ago
Tooth.Characterization.jpg	Only 'mode' changes	2 months ago
ToothAnalysis.ipynb	Only select a subset if we actually have data:wq	22 days ago
ToothDataSize.ipynb	Clean run of notebook	22 days ago
ToothDisplay.ipynb	Display Tooth045 for manuscript	22 days ago
requirements.txt	We also need this	2 months ago
treebeard.yaml	Add treebeard configuration	20 days ago

Below the file list, the README.md content is shown. It includes a DOI: 10.5281/zenodo.3999402, a 'treebeard.yml' file with a 'failing' status, and a 'launch binder' button. The title of the README is 'A big tooth cohort'. The text describes the cohort and provides instructions on how to access the analysis notebook via Binder.




u^b

μ CT imaging



Dataset cropping




- Full datasets: 326 GB
- Cropped datasets: 115 GB



u^b


Dataset cropping

- Full datasets: 326 GB
- Cropped datasets: 115 GB




u^b

Tooth morphology



u^b


Tooth morphology



From [25], Fig. 2

u^b


Tooth morphology




u^b

Detection of enamel-dentin border

Sagittal MIP
Slices and grey value profile




Coronal MIP
Slices and grey value profile




Tooth045


Minima of 40 slices around slice 950




Minima of 40 slices around slice 1434



Minima of 40 slices around slice 1918



Minima of 40 slices around slice 2351




u^b


Detection of enamel-dentin border

Tooth045


Tooth045, Slice 950: 1




Slice 1434: 2



Slice 1918: 2



Slice 2351: 1




Classification of root canal configurations

Roots		RCC	#	%
Single (N=98)		1-1-1/1	73	74.5
		1-1-1/2	14	14.3
		1-1-1/3	1	1.0
		1-1-1/4	2	2.1
		1-1-2/1	1	1.0
		1-2-1/1	4	4.1
		1-2-1/2	1	1.0
		1-2-2/2	1	1.0
		2-3-1/1	1	1.0
Double (N=3)	Buccal	1-1-1/1	2	66.6
		1-2-1/1	1	33.3
	Lingual	1-1-1/1	2	66.6
		1-1-1/2	1	33.3


u^b

Extraction of root canal space




u^b

Results of root canal space extraction



u^b


Physiological foramen geometry



From [27], Fig. 1

u^b

Physiological foramen geometry




Conclusion ZMK

- Efficient use of time, e. g. more teeth does not mean more (human) work
- Reproducible analysis with *free and open-source* software, usable by *anyone*
- Objective analysis, e. g. no operator bias

u^b

Metal foam




Etienne Berner NanoElectroCatalysis Group

Data wrangling by example: Cichlids

Collaboration with team of *Aquatic Ecology & Evolution*, from the Institute of Ecology and Evolution^a

- 133 Cichlids from Lake Victoria, East Africa
 - Functional anatomy of the skulls and jaws
 - 6–18 cm in size
- 375 scans in total
 - Voxelsizes from 3.5–50 μm
 - 46 days of scanning time
 - 9.8 TB of raw data
 - 1.5 TB/+1 000 000 reconstructions




DOI:gsst8t, Fig. 1

^a11.


u^b

Visualization of cichlid head




u^b

Visualization of cichlid head





u^b

Visualization of segmented pharyngeal jaw








u^b

Data wrangling by example: Cichlids



u^b

Data wrangling by example: Cichlids



u^b

Thanks!

- Thanks for listening to me!
- What questions do you have for me?

Colophon

- This BEAMER presentation was crafted in \LaTeX with the (slightly adapted) template from *Corporate Design und Vorlagen* of the University of Bern.
 - Complete source code: git.io/fjpP7
 - The \LaTeX code is automatically compiled with a GitHub action to a (handout) PDF which you can access here: git.io/JeQxO
- Did you spot an error?
 - File an issue: git.io/fjpPb
 - Submit a pull request: git.io/fjpPN
 - Send me an email: david.haberthuer@unibe.ch

References I

- [1] Ruslan Hlushchuk et al. "Cutting-Edge Microangio-CT: New Dimensions in Vascular Imaging and Kidney Morphometry". In: (Mar. 2018). DOI: 10.1152/ajprenal.00099.2017.
- [2] Henry Nording et al. "The C5a/C5a Receptor 1 Axis Controls Tissue Neovascularization through CXCL4 Release from Platelets". In: (Dec. 2021). DOI: 10.1038/s41467-021-23499-w.
- [3] Ruslan Hlushchuk et al. "Innovative High-Resolution microCT Imaging of Animal Brain Vasculature". In: (Oct. 2020). DOI: 10.1007/s00429-020-02158-8.
- [4] Tsering Wüthrich et al. "Development of Vascularized Nerve Scaffold Using Perfusion-Decellularization and Recellularization". In: (Aug. 2020). DOI: 10.1016/j.msec.2020.111311.
- [5] Cédric Zubler et al. "The Anatomical Reliability of the Superficial Circumflex Iliac Artery Perforator (SCIP) Flap". In: (Mar. 2021). DOI: 10.1016/j.aanat.2020.151624.
- [6] Matthias Messerli et al. "Adaptation Mechanism of the Adult Zebrafish Respiratory Organ to Endurance Training". In: (Feb. 2020). DOI: 10.1371/journal.pone.0228333.
- [7] Verdiana Trappetti et al. "Synchrotron Microbeam Radiotherapy for the Treatment of Lung Carcinoma: A Pre-Clinical Study". In: (Aug. 2021). DOI: 10.1016/j.ijrobp.2021.07.1717.
- [8] Estée Bochud et al. "A New Diancta Species of the Family Diplommatinidae (Cyclophoroidea) from Vanua Levu Island, Fiji". In: (Nov. 2021). DOI: 10.3897/zookeys.1073.73241.

References II

- [9] Sebastian Halm et al. "Micro-CT Imaging of Thiel-embalmed and Iodine-Stained Human Temporal Bone for 3D Modeling". In: (2021). DOI: 10.1186/s40463-021-00522-0.
- [10] David Haberthür et al. "Automated Segmentation and Description of the Internal Morphology of Human Permanent Teeth by Means of Micro-CT". In: (Apr. 2021). DOI: gjpw2d.
- [11] David Haberthür et al. "Microtomographic Investigation of a Large Corpus of Cichlids". In: (Sept. 2023). DOI: gsst8t.
- [12] A. M. Cormack. "Representation of a Function by Its Line Integrals, with Some Radiological Applications". In: (Sept. 1963). DOI: 10.1063/1.1729798.
- [13] Godfrey Newbold Hounsfield. "Historical Notes on Computerized Axial Tomography.". In: (1976).
- [14] E C Beckmann. "CT Scanning the Early Days.". In: (Jan. 2006). DOI: 10.1259/bjr/29444122.
- [15] J Hsieh. *Computed Tomography: Principles, Design, Artifacts, and Recent Advances*. Society of Photo Optical, 2003.
- [16] L. A. Feldkamp et al. "Practical Cone-Beam Algorithm". In: (June 1984). DOI: 10.1364/JOSAA.1.000612.
- [17] L. A. Feldkamp et al. "The Direct Examination of Three-Dimensional Bone Architecture in Vitro by Computed Tomography". In: (Feb. 1989). DOI: 10.1002/jbmr.5650040103.
- [18] M. W. Layton et al. "Examination of Subchondral Bone Architecture in Experimental Osteoarthritis by Microscopic Computed Axial Tomography". In: (Nov. 1988). DOI: 10.1002/art.1780311109.

References III

- [19] Mark Hammer. *X-Ray Physics: X-Ray Interaction with Matter, X-Ray Contrast, and Dose*.
- [20] Wikipedia contributors. *Beer–Lambert Law — Wikipedia, The Free Encyclopedia*. 2019.
- [21] Kenneth Clark et al. “The Cancer Imaging Archive (TCIA): Maintaining and Operating a Public Information Repository”. In: (Dec. 2013). DOI: 10.1007/s10278-013-9622-7.
- [22] Johannes Schindelin et al. “Fiji: An Open-Source Platform for Biological-Image Analysis”. In: (July 2012). DOI: 10.1038/nmeth.2019.
- [23] Thomas Kluyver et al. “Jupyter Notebooks – a Publishing Format for Reproducible Computational Workflows”. In: (2016). DOI: 10.3233/978-1-61499-649-1-87.
- [24] Thomas Gerhard Wolf et al. “Internal Morphology of 101 Mandibular Canines of a Swiss-German Population by Means of Micro-CT: An Ex Vivo Study”. In: (2021). DOI: g7r8.
- [25] Benjamín Briseño-Marroquín et al. “Root Canal Morphology and Configuration of 179 Maxillary First Molars by Means of Micro-Computed Tomography: An Ex Vivo Study”. In: (Dec. 2015). DOI: 10.1016/j.joen.2015.09.007.
- [26] David Haberthür. “Habi/Zmk-Tooth-Cohort: Used for Manuscript about Method”. In: (Aug. 2020). DOI: 10.5281/ZENODO.3999402.
- [27] Thomas Gerhard Wolf et al. “Three-Dimensional Analysis of the Physiological Foramen Geometry of Maxillary and Mandibular Molars by Means of Micro-CT”. In: (Sept. 2017). DOI: 10.1038/ijos.2017.29.